



Date created: 16 Oct 2020 15:44:11 GMT
Source: pogoltdemo@gmail.com

Date Ranges	From	To
Requested	14 Oct 2020	16 Oct 2020
Received	14 Oct 2020	16 Oct 2020

Reporting Performance	Totals
Number of reports received	7
Number of reports with 30DTR data	7
Number of unique loggers	5
Number of unique loggers with alarms	4
Number of manual reports (no 30DTR data)	0

Temperature Performance	Percentages
Percentage of measurements between 2 to 8°C	79.52
Percentage of measurements outside of 2 to 8°C (Alarm State)	19.4
Percentage of measurements outside of 2 to 8°C (No Alarms)	1.07
Percentage of measurements with data errors	0.01

30 DTR Reports

serial	alarm count	alarm time %	excursion time %	safe time %	data error %	facility	model	facility contact	contact phone number	notes
350600018158	60	96.58	0.7	2.72	0	Health Facility Alpha	Other	John Doe	8675309	
130400016731	3	0.94	2.86	96.2	0	Test 3	Other	John Smith	97576757	
160600006358	2	0.38	0.54	99.08	0	Health Facility Beta	Other	Jane Smith	6669696	
350600043859	1	0.38	1.19	98.43	0	Test 2	Other	Jane Doe	67873759464	
85V-0002	0	0	0	100	0	Facility 4	CFD-50 SDD	John Doe	1358976442	

Manual Reports

vx chamber cold	temperature	facility	lga/subcounty	location	contact	phone	notes
-----------------------	-------------	----------	---------------	----------	---------	-------	-------